

BPW SWAB

Presentation

Conical bottom plastic tube with screw cap and leak-proof rings containing 10ml and cap with sterile flocced nylon swab.

Sterilization method

Gamma irradiation.

Application

Used as a dragging swab for environmental monitoring, especially for collecting samples for the detection of *Salmonella* spp. The presence of Buffered Peptone Water (BPW) ensures non-selective pre-enrichment of microorganisms present in the collected samples, keeps the cells viable and promotes the recovery of injured cells until the laboratory analysis.

Principle

Non-selective enriched medium that allows the growth of various microorganisms, particularly members of pathogenic Enterobacteria such as *Salmonella*. It promotes the recovery of injured cells by incubating the sample under non-selective conditions for at least 18 hours.

How to use

Scrub the swab over the desired surface according to the adopted sampling plan. At the end of the collection, place the swab into the tube, break off the handle, and close the cap. Send the sample to the laboratory and proceed with the analysis according to the adopted methodology.

Quality Control

Test	Result
Sterility	Absence of microbial growth
<i>Salmonella enterica</i> ATCC 14028	Good growth with turbidity of the medium
<i>Escherichia coli</i> ATCC 25922	Good growth with turbidity of the medium
Appearance	Liquid medium, light yellow to very clear, clear, may contain slight precipitate
pH at 25°C	7.0 ± 0.2

Results interpretation

Microbial growth is evidenced by the turbidity of the medium. If growth is observed, perform microscopic analysis, subculture on selective media, and biochemical tests to identify isolated genera and species, if necessary. Perform the reading according to official compendia or internal laboratory methodology.

Precautions and special care

Product intended for *in vitro* diagnostic use only.

Restricted for use by professionals. Do not inhale or ingest.

Do not use the product beyond the expiration date, with signs of contamination, or if it has changed color. In the presence of contamination, the product should be immediately discarded.

Do not use the product if the packaging is damaged or tampered with.

Storage

Store between 2-35°C in a dry place and protect from light.

Shelf-life

1 year from the date of manufacture.

Disposal of the product

After use, the product must be handled at the generating unit before environmentally appropriate final disposal, in accordance with official regulations.

Quality Guarantee

bioBoaVista guarantees the quality of its products as long as they are used according to their respective instructions and in accordance with national and international references. bioBoaVista does not take responsibility for the use of its products for purposes other than those described and approved by the company. All clinical diagnoses should be analyzed in conjunction with clinical evidence and not solely based on laboratory results.

References

1. Becton, Dickinson and Company. Difco & BBL Manual. Manual of Microbiological Culture Media, 2nd ed., 2009.
2. ISO 6579-1:2017. Microbiology of food chain – Horizontal method for the detection, enumeration and serotyping of *Salmonella*. Part 1: Detection of *Salmonella* spp.
3. ISO 11133:2014. Microbiology of food, animal feed and water - Preparation, production, storage and performance testing of culture media.
4. Manual de Métodos de Análise Microbiológica de Alimentos, Livraria Varela, 3ª ed., 2007.
5. Merck Microbiology Manual. 12th ed.